BIOBOON BIOTECH LLC

OPIATE (MORPHINE) RAPID SCREENING TEST KIT

Rapid Drugs (Opiate / Morphine) Screening Test Kit (Urine)





SPECIMEN: Urine

CUT-OFF: OPI 300 ng/mL

RESULT TIME: 5 Minutes

INTERPRETATION OF RESULTS



POSITIVE

The control line is the only visible line on the test device. This is indicative of presence of OPI above 300 ng/mL.



NEGATIVE

The control line and Test line are visible on the test device. This no OPI detected or OPI below 300 ng/mL.



INVALID

The control line fails to appear.

DIRECTIONS FOR USE

1. Allow the test device, specimen and/or buffer to equilibrate at room temperature (15–30°C) before testing.
2. Bring the pouch to room temperature before opening it. Remove the test cassette from the sealed pouch and use it within 1 hour.
3. Place the cassette on a clean and level surface.
4. Hold the dropper vertically and transfer 2 drops of urine (approximately 50 µL).
5. Wait for the colored line(s) to appear. Read results at 5 minutes.

Note: Do not interpret the result after 10 minutes.

PERFORMANCE

Relative Sensitivity:	100%
Relative Specificity:	98.57%



INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan
Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com