BIOBOON BIOTECH LLC

BARBITURATE RAPID TEST KIT

Rapid Drugs (Barbiturate) Screening Test Kit (Urine)





DIRECTIONS FOR USE

Allow the test device, specimen and/or buffer to equilibrate at room temperature (15–30°C) before testing.

1.  Bring the pouch to room temperature before opening it. Remove the test cassette from the sealed pouch and use it within 1 hour.
2.  Place the cassette on a clean and level surface.
3.  Hold the dropper vertically and transfer 2 drops of Urine (approximately 50 µL).
4.  Wait for the colored line(s) to appear. Read results at 5 minutes.
5. Do not interpret the result after 10 minutes.

INTERPRETATION OF RESULTS



POSITIVE

The control line is the only visible line on the test device. This is indicative of presence of BAR above 300 ng/mL.

NEGATIVE

The control line and Test line are visible on the test device. This no BAR detected or BAR below 300 ng/mL.

INVALID

The control line fails to appear.

PERFORMANCE

Relative Sensitivity:	100%
Relative Specificity:	99.28%

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan
Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com