BIOBOON BIOTECH LLC

Creatinine Test Kit (Enzymatic Method)



- Accurate & Reliable
- Enzymatic Method
- Wide Linear Range (up to 25 mg/dL)
- Reagent Ready to Use



TEST PARAMETERS

Wavelength	546 nm
Reaction Temperature	37°C
Measurement	Against Reagent
Reaction Type	End Point
Reaction Direction	Increasing
Sample Volume	20 µL
Reagent Volume	600 µL + 200 µL
Incubation	5 minutes
Linearity	25 mg/dL

MANUAL ASSAY PROCEDURE

Pipette into Test Tubes.

	Blank	Standard	Sample
Reagent 1	600 µL	600 µL	600 µL
Sample	-	20 µL	20 µL
Mix well and incubate for 5 min at 37°C and add			
Reagent 2	200 µL	200 µL	200 µL

Mix well and incubate for 5 min at 37°C. Measure the absorbance of sample and standard against reagent blank.

SAMPLE DILUTIONS

1. This method is linear upto a concentration of 25 mg/dL.
2. Dilute samples above this concentration 1:1 with 0.9% saline.
3. Repeat assay. Multiply the result by 2.

LINEARITY & RANGE

LINEARITY
The method is linear upto a concentration of 25 mg/dL.

RECOMMENDATION
Dilute samples above 25 mg/dL, 1:1 with 0.9% saline and multiply the result by 2.

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan
Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com