BIOBOON BIOTECH LLC

Dengue NS1 Ag Rapid Test Kit

Serum / Plasma / Whole Blood

- Rapid Antigen Detection
- Simple Test Procedure
- Read at 15 Minutes
- For In Vitro Diagnostic Use



PRINCIPLE

Immunochromatographic rapid test for qualitative detection of Dengue NS1 antigen in human serum, plasma or whole blood.

DIRECTIONS FOR USE

1. Bring test device, specimen and assay buffer to room temperature (15–30°C).
2. Remove device from pouch and use as soon as possible, preferably within 1 hour.
3. Place device on a clean, flat surface.
4. Add specimen to sample well (S) using the provided dropper, then add assay buffer to the sample well.
5. Read at 15 minutes; do not interpret after 20 minutes.

RESULT INTERPRETATION



POSITIVE	NEGATIVE	INVALID
Control line (C) and test line (T) visible.	Only control line (C) visible.	Control line (C) absent. Repeat with a new device.

PERFORMANCE

Relative Sensitivity:	98%
Relative Specificity:	100%
Overall Agreement:	98.57%

For in vitro diagnostic use only, Follow the product instructions.

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207, Regent Square
2nd floor, Above D'mart shopping mall
Adajan
Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com