BIOBOON BIOTECH LLC

FERRITIN RAPID TEST KIT

Serum / Plasma / Whole Blood





Read results within **15 minutes**

Do not interpret after **20 minutes**

ASSAY FEATURES

- Rapid ferritin detection
- Sensitivity: 98.9%
- Specificity: 98.8%

DIRECTIONS FOR USE

1. Allow test device, sample buffer and specimen to reach room temperature (15–30°C).
2. Open pouch and place device on a clean, flat surface.
3. Add 1 drop (approximately 25 µL) of serum, plasma or whole blood to the sample well.
4. Add 1 drop (approximately 40 µL) of sample buffer to the sample well.
5. Start timer; read within 15 minutes; do not interpret after 20 minutes.

RESULT INTERPRETATION



Positive	Negative	Invalid
Two red bands at C and T	One red band at C only	No red band at C, repeat with a new device.

Normal Ferritin Levels (ng/mL)

Adult males	24–336
Adult females	24–307
Newborns	25–200
Infants aged 1 month	200–600

For *in vitro* diagnostic use.

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan
Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com