BIOBOON BIOTECH LLC

hCG – Pregnancy Rapid Test Strip



Urine Test

- Early & Accurate Pregnancy Detection
- Rapid Results (3–5 Minutes)
- Simple & Easy to Use
- High Sensitivity (25 mIU/mL)





INTERPRETATION OF RESULTS



POSITIVE

In addition to the colored line in the control region, a clearly distinguishable pink line appears.



NEGATIVE

Only one pink colored line appears in the Control areaa.



INVALID

If no line appars in the control as well as the test region, the test should be repeated with fresh test strip.

DIRECTIONS FOR USE

1. Allow test strip and urine samples to equilibrate to room temperature (20–30°C) prior to testing.
2. Remove the hCG-Pregnancy Rapid Test Strip from pouch. Use strip as soon as possible but within 1 hour after removal from pouch especially if the room temperature is more than 30°C and in high humidity environment.
3. Immerse the test strip in the urine sample with printed sample end pointing toward the urine for at least 5 seconds. Be sure the sample level is below the marked sample line on test strip.
4. Wait for red bands to appear. The test should be read in approximately 3–5 minutes for urine.

PERFORMANCE

SENSITIVITY

The analytical sensitivity of the hCG-Pregnancy Rapid Test Strip is 25 mIU/mL.

SPECIFICITY

Negative results were obtained from all tests conducted with 500 mIU/mL hLH, 1000 mIU/mL hFSH and 1000 µIU/mL hTSH.

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan, Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com