BIOBOON BIOTECH LLC



JAPANESE ENCEPHALITIS IgM Ab RAPID TEST

Serum / Plasma / Whole Blood

- Rapid results — read at 15 minutes
- Serum / Plasma / Whole Blood
- Simple 4-step procedure
- For in vitro diagnostic use only



RESULT INTERPRETATION



POSITIVE | NEGATIVE | INVALID

TEST PROCEDURE

1. Bring pouch, specimen and buffer to room temperature (15–30°C).
2. Open pouch and use cassette within 1 hour; place on a clean, level surface.
3. Add 1 drop of serum/plasma (approx. 25 µL) to the specimen well, then 1 drop of buffer (approx. 40 µL).
4. Read at 15 minutes; do not interpret after 20 minutes.

PERFORMANCE Relative sensitivity: 97.3% | Relative specificity: 99%

Interpret results with clinical findings.

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan, Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063, USA
Email: mahipal@bioboonbiotech.com