BIOBOON BIOTECH LLC



Luteinizing Hormone (LH) Rapid Test Kit

Cassette / Midstream • Urine

- **Urine Sample**
- **Easy to Use**
- **Read at 5 Minutes**





RESULT INTERPRETATION



POSITIVE

Two lines are visible and the Test (T) line is as dark as or darker than the Control (C) line.



NEGATIVE

Two lines are visible, but the Test (T) line is lighter than the Control (C) line, or there is no Test (T) line.



INVALID

The control line does not appear.

DIRECTIONS FOR USE

1. Allow test, urine specimen and controls to reach room temperature (15–30°C).
2. Bring the pouch to room temperature, open and use cassette promptly.
3. Place the cassette on a clean, level surface.
4. Add 3 full drops of urine (approximately 100 µL) to the sample well (S).
5. Read at 5 minutes. Do not interpret after 10 minutes.

PERFORMANCE

Relative Sensitivity: >99.9% (98.7%–100%)

Relative Specificity: >99.9% (99.2%–100%)

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan, Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com