BIOBOON BIOTECH LLC

Innovative Diagnostics for a Healthier Tomorrow



NT-ProBNP Rapid Test Kit

Serum / Plasma / Whole Blood

Rapid Cardiac Marker Screening

Qualitative detection of NT-ProBNP in human serum, plasma or whole blood.

For In-Vitro Diagnostic Use Only.



KEY BENEFITS

- **Early Screening**
- **Simple Procedure**
- **Clear Visual Results**

SAMPLE TYPES



- **Whole Blood** (venipuncture or fingerstick)
- **Serum** (after clotting)
- **Plasma** (with anticoagulant such as heparin, EDTA or sodium citrate)

TEST PROCEDURE

1. Bring sealed pouch to room temperature (15–30°C).



2. Place the cassette on a clean, level surface.



3. Add specimen and buffer according to the kit instructions.



4. Read results at 15 minutes. Do not interpret results after 20 minutes.



RESULT INTERPRETATION

POSITIVE



C and T lines visible.

NEGATIVE



C line only.

INVALID



No C line.
Repeat with a new device.

STORAGE

- Store at 2–30°C (room temperature).
- Do not freeze.
- Keep sealed pouch until use.
- Do not use if the pouch is damaged or the kit is expired.

India Office
BIOBOON BIOTECH INDIA LLP
207, Regent Square, 2nd Floor,
Above D'Mart Shopping Mall, Adajan, Surat - 395009
info@bioboonbiotech.com

Global Diagnostic Solutions for a Healthier Tomorrow

USA Address
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ 07063, USA
mahipal@bioboonbiotech.com