BIOBOON BIOTECH LLC



Scrub Typhus IgG/igM Ab Rapid Test Kit

Serum / Plasma / Whole Blood

- IgG & IgM Antibody Detection
- Read at 15 Minutes
- Simple Test Procedure
- For In Vitro Diagnostic Use



RESULT INTERPRETATION



IgG POSITIVE	IgM POSITIVE	IgG/IgM POSITIVE	NEGATIVE	INVALID

DIRECTIONS FOR USE

1. Allow device, specimen and/or buffer to reach room temperature (15–30°C).
2. Bring pouch to room temperature; open and use cassette within 1 hour.
3. Place cassette on a clean, level surface.
4. For serum/plasma, add 1 drop (approximately 25 µL) to the specimen well, then add 1 drop of buffer (approximately 40 µL).
5. Read at 15 minutes. Do not interpret after 20 minutes.

PERFORMANCE

- IgM Relative Sensitivity: 100%
- IgG Relative Sensitivity: 100%
- Relative Specificity: 99.11%

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan
Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com