BIOBOON BIOTECH LLC

SYPHILIS RAPID TEST KIT

Serum / Plasma / Whole Blood





DIRECTIONS FOR USE

Allow test device, assay buffer and specimen to equilibrate to room temperature (15–30°C) before testing.

1.  Add 1 drop (25 µL) of serum/plasma/whole blood to sample well "S".
2.  Add 1 drop (40 µL) of assay buffer to sample well "S".
3. Allow reaction to occur for 15–20 minutes.
4.  Read the test between 15–20 minutes after adding specimen.

RESULT INTERPRETATION



Positive

In addition to the colored line in the control region (C), a clearly distinguishable pink purple colored line also appears in the test region (T), indicating a positive result.



Negative

Only one pink-purple colored line appears at the control region (C). The test result is negative.



Invalid

If no line appears in the control region (C), the test is invalid. Repeat the test with a fresh cassette/test.

Warning: Do not interpret results after 20 minutes.

INDIA OFFICE
BIOBOON BIOTECH INDIA LLP
207. Regent Square
2nd floor, Above D'mart shopping mall
Adajan
Surat-395009
Email: info@bioboonbiotech.com

USA ADDRESS
BIOBOON BIOTECH LLC
1526 West 3rd Street,
Plainfield, NJ, 07063
USA
Email: mahipal@bioboonbiotech.com